

PEAK SHAPE PERFORMANCE



UHPLC PERFORMANCE
WITH ANY LC INSTRUMENT

SUNSHELL

HARDCORE SHELL
TECHNOLOGY

ChromaNyk
ChromaNik Technologies Inc.

GLOBAL DISTRIBUTOR

BIOTECH AB

www.biotech.se



PEAK SHAPE
PERFORMANCE



HARDCORE



SUNSHELL



SUPERFICIALLY
POROUS SILICA



EFFICIENT

WHAT IS SUNSHELL?



THE NEXT GENERATION HARDCORE SHELL PARTICLE

Secure your analysis with SunShell hardcore column technology

Unique bonding technology combined with core shell particles gives you faster performance and more reliable results. The SunShell technique assures top efficiency with all kinds of LC and UHPLC systems.

FEATURES OF SUNSHELL 2.6 μm AND 5 μm


- . 1.6 μm and 3.4 μm of core and 0.5 μm and 0.6 μm of superficially porous silica layer.
- . Same efficiency and high throughput as a Sub-2 μm and 3 μm particle.
- . Same pressure as a 3 μm and 5 μm particles.
- . Same chemistry as Sunniest technology (reference figure below).
- . Good peak shape for all compounds such as basic, acidic and chelating compounds.
- . High stability (pH range for SunShell C18, 1.5 to 10).
- . Low bleeding.



Schematic diagram of bonding of SunShell C18

SunShell C18 shows same efficiency as a Sub 2 μm C18. In comparison between fully porous 2.6 μm and core shell 2.6 μm (SunShell), SunShell shows lower values for A term, B term and C term of Van Deemter equation. The core shell structure leads to higher performance compared with the fully porous structure.

Furthermore back pressure of SunShell C18 is less than a half compared to Sub-2 μm C18s.



Comparison of back pressure for high throughput columns



HOW DOES SUNSHELL WORK?


NARROW PARTICLE DISTRIBUTION

VAN DEEMTER EQUATION

Van Deemter Equation

$$H = A d_p + B \frac{D_m}{u} + C \frac{d_p^2}{D_m} u$$

A term : Eddy diffusion (d_p is particle diameter)
B term : Longitudinal diffusion
(D_m is diffusion coefficient)
C term : Mass transfer




A TERM


The size distribution of a core shell (SunShell) particle is much narrower than that of a conventional totally

porous particle, so that the space in between the particles in the column is reduced and efficiency increases by

reducing Eddy Diffusion (multi-path diffusion) as the A term in Van Deemter Equation.



Wide particle distribution
(Conventional silica gel $D_{90}/D_{10}=1.50$)




Narrow particle distribution
(Core Shell silica $D_{90}/D_{10}=1.15$)


B TERM

Diffusion of a solute is blocked by the existence of a core, so that a solute diffuses less in a core shell silica column than in a totally porous silica column. Consequently B term in Van Deemter Equation reduces in the core shell silica column.


Totally porous silica A solute diffuses in a pore as well as outside of particles.



Core shell silica A core without pores blocks diffusion of a solute.




Difference in longitudinal diffusion




Plot of Plate height vs Flow rate

C TERM



As shown in the left figure, a core shell particle has a core so that the diffusion path of samples shortens and mass transfer becomes fast. This means that the C term in Van Deemter Equation reduces. In other words, HETP (theoretical plate) is kept even if flow rate increases. A 2.6 μm core shell particle shows the same column efficiency as a totally porous Sub-2 μm particle.

The right figure shows the diffusion width of a sample in a 2.6 μm core shell particle and a 2 μm totally porous particle. Both diffusion widths are almost the same. The 2.6 μm core shell particle is superficially porous, so that the diffusion width becomes narrower than particle size. Same diffusion means same efficiency.



Diffusion of a sample in core shell and totally porous silica




C18 - 2.6 μm

HIGHEST RETENTION / LARGEST STERIC SELECTIVITY / LOWEST BACKPRESSURE

Retention of standard samples and back pressure were compared for five kinds of core shell type C18s. Company A C18 showed only a half retention in comparison with SunShell C18. Steric selectivity becomes large when ligand density on the surface is high. SunShell C18 has the largest steric selectivity as well as the highest ligand density leading to the longest retention time.

SUNSHELL C18 COMPARISON



Mobile phase: CH₃OH/H₂O=75/25
Flow rate: 1.0 mL/min, Temperature: 40° C
Sample: 1 = Uracil, 2 = Caffeine, 3 = Phenol,
4 = Butylbenzene, 5 = o-Terphenyl,
6 = Amylbenzene, 7 = Triphenylene


	Hydrogen bonding	Hydrophobicity	Steric selectivity
Company A C18	0.48	1.54	1.20
Company B C18	0.35	1.56	1.50
Company C C18	0.42	1.57	1.25
Company D C18	0.44	1.60	1.31
Sunshell C18	0.39	1.60	1.46

BEST PEAK SHAPE AVAILABLE

Amitriptyline overloads much more at acetonitrile/buffer mobile phase than methanol/buffer which causes tailing. Five kinds of core shell C18s were


compared as refers to loading capacity of amitriptyline. Thanks to the unique bonding technology Sunshell gives extraordinary peak shape, which means

better sensitivity and accuracy of the method.



Mobile Phase:

• Acetonitrile/20 mM phosphate buffer pH 7.0 (60/40)



Mobile Phase:

• Acetonitrile/10 mM ammonium acetate pH 6.8 (40/60)

Company A C18: Kinetex C18


Company B C18: Accucore C18

Company C C18: PoroShell C18 EC

Company D C18: Ascentis Express C18

EXPANDED pH RANGE DUE TO THE SUNSHELL BONDING TECHNOLOGY


SUNSHELL C18 STABILITY



Durable test condition

Column size: 50 x 2.1 mm
 Mobile phase: CH₃CN/1.0% TFA, pH1=10/90
 Flow rate: 0.4 mL/min
 Temperature: 80 °C

Stability under acidic pH condition was evaluated at 80°C using acetonitrile/1% trifluoroacetic acid solution (10:90) as mobile phase. 100% aqueous mobile phase expels from the pores of C18 packing materials by capillarity and packing materials do not deteriorate. Adding 10% acetonitrile to the mobile phase enables accurate evaluation.



Durable test condition


Column Size: 50 x 2.1 mm
 Mobile phase:
 CH₃OH/20mM Sodium borate/10mM NaOH=30/21/49 (pH10)
 Flow rate: 0.4 mL/min
 Temperature: 50 °C

Stability under basic pH condition was evaluated at 50°C using methanol/Sodium borate buffer pH 10 (30:70) as mobile phase. Sodium borate is used as an alkaline standard solution for pH meters, which allows for a high buffer capacity. Elevated temperature of 10°C reduces column life to one third. The other company shows stability when tested at ambient (room) temperature. If room temperature is 25°C, column life is sixteen times longer than at 50°C.

BLEEDING TEST USING LC/MS

.....

The high stability of the SunShell columns also means low bleeding in LC/MS analysis as shown here.



Column size: 50 x 2.1 mm

Mobile phase:

A) 0.1% Acetic acid

B) CH₃CN

Gradient:

Time: 0min 1min 5min 7min

%B: 5% 5% 100% 100%

Flow rate: 0.4 mL/min

Temperature: 40 °C

MS: ABI API-4000 Ionization:


Turboionspray (cation)

Measurement mode:

Q1 Scan m/z 100-1000

SUNSHELL C18 EFFICIENCY

Column: SunShell C18, 2.6 μ m 50 x 2.1 mm



Column: SunShell C18, 2.6 μ m 50 x 2.1 mm

Mobile phase: CH₃CN/H₂O=60/40

Flow rate: 0.3 mL/min

Pressure: 7 MPa

Temperature: 23 °C

UHPLC: Jasco X-LC


Sample: 1 = Uracil

2 = Toluene

3 = Acenaphthene

4 = Butylbenzene

Column: SunShell C18, 2.6 μ m 150 x 4.6 mm



Column: SunShell C18, 2.6 μ m 150 x 4.6 mm

Mobile phase: CH₃CN/H₂O=70/30

Flow rate: 1.0 mL/min

Pressure: 15.5MPa

Temperature: 25 °C

UHPLC: Jasco X-LC


EFFICIENCY = 253,000 plates/m

ORDERING INFO OF SUNSHELL	Inner diameter (mm)	1.0	2.1	3.0	4.6	USP category
	Length (mm)	Catalog no	Catalog no	Catalog no	Catalog no	Catalog no
Sunshell C18, 2.6 μ m	30	---	CB6931	CB6331	CB6431	L1
	50	CB6141	CB6941	CB6341	CB6441	
	75	---	CB6951	CB6351	CB6451	
	100	CB6161	CB6961	CB6361	CB6461	
	150	CB6171	CB6971	CB6371	CB6471	
	250	---	---	CB6381	CB6481	



C18 - 5 μm

Can be used in any LI method - but with improved performance.



Column size: 150 x 4.6 mm
Mobile phase: CH₃OH/H₂O=75/25
Flow rate: 1.0 mL/min
Temperature: 40° C

Sample: 1 = Uracil

2 = Caffeine

3 = Phenol

4 = Butylbenzene

5 = o-Terphenyl

6 = Amylbenzene

7 = Triphenylene

HPLC: Hitachi LaChrom ELITE (Tubing, 0.25 mm i.d.)



	Totally porous silica Sunniest C18, 5 μm		Core shell silica SunShell C18, 2.6 μm		Core shell silica SunShell C18, 5 μm	
Specific surface area	340 m ² /g		150 m ² /g		90 m ² /g	
	Retention time (t_R)	Retention factor (R)	Retention time (t_R)	Retention factor (R)	Retention time (t_R)	Retention factor (R)
1) Uracil	1.70	0	1.34	0	1.30	0
6) Amylbenzene	19.96	10.74	16.56	11.36	13.43	9.33
Relative value of Amylbenzene	100%	100%	83%	106%	67%	87%


There is a small difference of R between totally porous and core shell particles.

ORDERING INFO OF SUNSHELL	Inner diameter (mm)	1.0	2.1	3.0	4.6	USP category
	Length (mm)	Catalog no	Catalog no	Catalog no	Catalog no	Catalog no
Sunshell C18, 5 μm	150	---	---	CB3371	CB3471	LI
	250	---	---	CB3381	CB3481	



ULTIMATE SELECTIVITY FOR YOUR ANALYSIS

C18 -WP / RP-AQUA / C8 / PHENYL / PFP - 2.6 μm



Column: SunShell C18, C18-WP, RP-AQUA, C8, Phenyl, PFP, 2.6 μm

150 x 4.6 mm

Mobile phase: CH₃OH/H₂O=75/25

Flow rate: 1.0 mL/min

Temperature: 40° C

Sample: 1 = Uracil

2 = Caffeine


3 = Phenol

4 = Butylbenzene

5 = o-Terphenyl

6 = Amylbenzene

7 = Triphenylene



	Hydrogen bonding	Hydrophobicity	Steric selectivity
PFP	1.00	1.31	2.38
Phenyl	1.00	1.48	1.01
C8	0.32	1.46	1.08
RP-AQUA	0.52	1.52	1.30
C18-WP	0.40	1.55	1.35
Sunshell C18	0.39	1.60	1.46

C18-WP / RP-AQUA / C8 / PHENYL / PFP - 2.6 µm

.....

ORDERING INFO OF SUNSHELL	Inner diameter (mm)	1.0	2.1	3.0	4.6	USP category
	Length (mm)	Catalog no	Catalog no	Catalog no	Catalog no	Catalog no
Sunshell C8, 2.6 µm	30	---	CC693I	CC633I	CC643I	L7
	50	---	CC694I	CC634I	CC644I	
	75	---	CC695I	CC635I	CC645I	
	100	---	CC696I	CC636I	CC646I	
	150	---	CC697I	CC637I	CC647I	
Sunshell PFP, 2.6 µm	30	---	CF693I	CF633I	CF643I	L43
	50	---	CF694I	CF634I	CF644I	
	75	---	CF695I	CF635I	CF645I	
	100	---	CF696I	CF636I	CF646I	
	150	---	CF697I	CF637I	CF647I	
Sunshell C18-WP, 2.6 µm	30	---	CW693I	CW633I	CW643I	L1
	50	---	CW694I	CW634I	CW644I	
	75	---	CW695I	CW635I	CW645I	
	100	---	CW696I	CW636I	CW646I	
	150	---	CW697I	CW637I	CW647I	
Sunshell RP-AQUA, 2.6 µm	30	---	CR693I	CR633I	CR643I	Equivalent to L62
	50	CR614I	CR694I	CR634I	CR644I	
	75	---	CR695I	CR635I	CR645I	
	100	CR616I	CR696I	CR636I	CR646I	
	150	CR617I	CR697I	CR637I	CR647I	
Sunshell Phenyl, 2.6 µm	30	---	CP693I	CP633I	CP643I	L11
	50	---	CP694I	CP634I	CP644I	
	75	---	CP695I	CP635I	CP645I	
	100	---	CP696I	CP636I	CP646I	
	150	---	CP697I	CP637I	CP647I	




HFC18 - 16 / HFC18 - 30 - 2.6 μm

High speed separations of proteins and peptides.


What is HFC18? Hexa-Functional C18 has six functional groups.

The HFC18 is much more stable under acidic conditions.




Proteins/peptides are often analysed at acidic pH. The wide pore SunShell phases are optimized for superior life time at extreme conditions.

HFC18 - 16 / HFC18 - 30 - 2.6 μm



Pore distribution of core shell particle



Stability under LC/MS mobile phase condition

Durable test condition

Column : SunShell HFC18-16 2.6 μm, 50 x 2.1 mm

Mobile phase:

CH₃CN/0.1% formic acid, pH2.6=40/60

Flow rate: 0.4 mL/min

Temperature: 70 °C

Measurement condition

Mobile phase: CH₃CN/H₂O=60/40

Flow rate: 0.4 mL/min

Temperature: 40 °C

Sample: 1 = Uracil

2 = Butylbenzene

ORDERING INFO OF SUNSHELL	Inner diameter (mm)	I.O	2.1	3.0	4.6	USP category
	Length (mm)	Catalog no	Catalog no	Catalog no	Catalog no	Catalog no
Sunshell HFC18-16, 2.6 μm	50	---	CG6941	CG6341	CG6441	L1
	100	---	CG6961	CB6361	CG6461	
	150	---	CG6971	CB6371	CB6471	
Sunshell HFC18-30, 2.6 μm	50	---	C46941	C46341	C46441	L1
	100	---	C46961	C46361	C46461	
	150	---	C46971	C46371	C46471	



HARDCORE SFC SEPARATIONS

2-EP (ETHYLPYRIDINE) - 2.6 μm

The 2.6 μm core shell column shows only one third of back pressure in comparison with the 1.7 μm fully porous column. However, both show almost

the same efficiency. By such low back pressure, a difference of density of supercritical fluid between an inlet and an outlet of the column is reduced. Conse-

quently, 2.6 μm core shell column performs a superior separation for SFC.




Figure 1: Chromatogram of the separation for the 17-components mix using the Sun Shell 2-EP 150 x 3.0 mm column. A methanol gradient of < 2 minutes was used on the Agilent 1260 Infinity SFC system. SFC conditions: flow rate: 4.0mL/min; outlet pressure 160 bar; column temperature 55°C. Gradient program: 5.0-7.5% in 0.20 min, then 7.5-20% in 1.3 min and held at 20% for 0.2 min.

2-EP - 2.6 μm




Figure 2: Chromatogram of the separation for the 17-components mix using Acquity UPC2 Viridis 2-EP 100 x 3.0 mm column. 16 of the 17 components were resolved. A methanol gradient of < 2 minutes was used on the Agilent 1260 Infinity SFC system. SFC conditions: flow rate 3.5 mL/min; outlet pressure 160 bar; and column temperature 70°C. Gradient program: 5.0-12.5% in 1.0 min, 12.5% for 0.25 min, then 12.5-20% in 0.75 min. Courtesy of Pfizer Inc.

ORDERING INFO OF SUNSHELL	Inner diameter (mm)	1.0	2.1	3.0	4.6	USP category
	Length (mm)	Catalog no	Catalog no	Catalog no	Catalog no	Catalog no
Sunshell 2-EP, 2.6 μm	30	---	CE693I	CE633I	CE643I	
	50	---	CE694I	CE634I	CE644I	
	75	---	CE695I	CE635I	CE645I	
	100	---	CE696I	CE636I	CE646I	
	150	---	CE697I	CE637I	CE647I	




HILIC -AMIDE - 2.6 μm

For Hydrophilic Interaction Chromatography.


Highly efficient separation of very polar compounds. Rapid equilibration.

Stationary phase of HILIC-Amide



R: Hydrophilic group

The stationary phase of SunShell HILIC-Amide consists of AMIDE and HYDROPHILIC GROUP, so that this stationary phase is more polar than an individual group. High speed separation is a result of core shell structure that derives high efficiency and fast equilibration.



Column:

SunShell HILIC-Amide, 2.6 μm 100 x 4.6 mm,

Coreshell polyol, 2.7 μm 100 x 4.6 mm,

Core shell Silica, 2.7 μm 100 x 4.6 mm

Mobile phase:

Acetonitrile/20 mM ammonium acetate(pH4.7) = 8/2

Flow rate: 1.0 mL/min

Temperature: 40 $^{\circ}\text{C}$

Detection: UV@250 nm

Sample: 1 = Thymine, 2 = Uracil, 3 = Uridine, 4 = Cytosine, 5 = Cytidine


Regarding retention of cytidine, SunShell HILIC-Amide showed 30% higher retention factor than S core shell polyol.

ORDERING INFO OF SUNSHELL	Inner diameter (mm)	1.0	2.1	3.0	4.6	USP category
	Length (mm)	Catalog no	Catalog no	Catalog no	Catalog no	Catalog no
Sunshell HILIC-Amide, 2.6 μm	30	---	CH6931	CH6331	CH6431	L68
	50	---	CH6941	CH6341	CH6441	
	75	---	CH6951	CH6351	CH6451	
	100	---	CH6961	CH6361	CH6461	
	150	---	CH6971	CH6371	CH6471	



PROTECT YOUR COLUMNS


RP GUARD FILTER




- The filter is made of porous glass sized 4 mm i.d. and 4 mm thickness
- Pore size is 2 μ m
- Low dead volume structure
- Back pressure on glass filter is ca 0.1 MPa at 1.0 mL/min of flow rate
- Upper pressure limit is more than 60 MPa
- Available for 2.1 mm i.d to 4.6 mm i.d. columns

SunShell C18, 2.6 μ m 50 x 2.1 mm

Without Guard Filter



With Guard Filter




Mobile phase:
CH₃CN/H₂O=60/40 for 2.1 mm i.d.
CH₃CN/H₂O=70/30 for 4.6 mm i.d.

Flow rate:
0.3 mL/min for 2.1 mm i.d.
1.8 mL/min for 4.6 mm i.d.


Temperature: 25 °C
Detection: UV@250nm
Sample: 1 = Uracil
2 = Toluene
3 = Acenaphthene
4 = Butylbenzene

SunShell C18, 2.6 μ m 150 x 4.6 mm

Without Guard Filter



With Guard Filter



ORDERING INFO OF SUNSHELL RP GUARD FILTER

(available as a guard column for reversed phase because of C18 bonding)

No

Sunshell RP Guard Filter Starter Kit (holder, cartridge, tubing)

CGGAAKN

Sunshell RP Guard Filter for exchange (5 pcs)

CBGAAC

Sunshell RP Guard Filter holder

CBGAAH



GLOBAL DISTRIBUTOR

BIOTECH AB

Råövägen 300, SE-439 92 Onsala, Sweden


TEL: +46 (0)300 56 91 80

info@biotech.se www.biotech.se

View from our office windows



LOCAL DISTRIBUTOR



AUTHORIZED DISTRIBUTOR

MZ-Analysentechnik GmbH, Barcelona-Allee 17 • D-55129 Mainz

Tel +49 6131 880 96-0, Fax +49 6131 880 96-20

e-mail: info@mz-at.de, www.mz-at.de

