

Chemicals and Industrial Polymers

HPLC Analysis of Acrylamide and Acrylic Acid on
Discovery® HS F5

column Discovery HS F5, 15 cm × 4.6 mm I.D., 5 µm particles (567516-U)
mobile phase 0.05% formic acid
flow rate 1.0 mL/min
column temp. ambient
detector UV, 206 nm
injection 1 µL
sample 1.0 mg/mL each in mobile phase
Application No. [G002569](#)

