

IC-H and IC-H Plus Devices

Maxi-Clean™ and Extract-Clean™ IC devices are solid-phase extraction devices used to eliminate matrix interferences from samples prior to analyses by ion chromatography. Each device consists of either 0.5mL or 1.5mL of polystyrene-based packing sandwiched between polyethylene frits within an injection-molded medical-grade polypropylene housing (Figure 1). Samples and wash solvents are passed through the packing using a luer hub syringe (Maxi-Clean™ device) or a vacuum manifold (Extract-Clean™ device). As sample comes in contact with the packing, specific chemical interactions take place that selectively retain certain components of the matrix in the device while the remaining components pass through the device outlet. The chemical characteristics of the packing dictate which sample components are retained. Seven chemistries are currently available.

Successful application of IC devices requires:

- 1) proper conditioning of the device prior to sample application,
- 2) application of the sample at a rate slow enough to allow the chemical interaction to take place while the sample is in the device, and
- 3) control of sample size to keep within the device's capacity.

The following information provides general recommendations for the use of IC-H devices. This procedure may be modified to accommodate samples with different characteristics.

General Information

Maxi-Clean™ and Extract-Clean™ IC-H devices provide a reliable method for the adjustment of pH of basic samples prior to the analysis of anions by ion chromatography. They may also be useful for removing cations from samples prior to the analysis of anions. IC-H devices contain either 0.5mL or 1.5mL of sulfonic acid cation exchange resin in the H⁺ form. The hydronium contained on the packing reacts with hydroxide from the sample to form water thereby reducing pH. In this process, cations from the sample are taken up by the resin to replace the hydronium consumed in the neutralization reaction. The net result is removal of hydroxide (and an equivalent volume of sample cations) from the matrix.

Polyvalent cations will be preferentially retained over monovalent cations in acidic samples providing a mechanism for removal of polyvalent cations.

The recovery of weakly ionized anions such as nitrite and phosphate could be poor after sample pretreatment with IC-H. The loss of these anions may take place through a number of different mechanisms, including evolution of nitrous oxide, oxidation of nitrite to nitrate, adsorption in the resin or a combination of effects. A study of the recovery of ions of interest in each sample matrix should be determined before quantitative results are obtained.




Figure 1

Flow Rate

The devices have a number of flow-dependent parameters that may affect results. In general, high flow rates, particularly in the sample loading step, will decrease the performance while low flow rates will improve the extraction process. Low flow rates allow the sample to diffuse into the packing thus increasing capacity and improving the efficiency. The recommended flow rate for sample loading is 1mL/minute or less.

Sample Mass

Each device contains either 0.8 or 2.0 milliequivalents of H⁺ (for the 0.5mL or 1.5mL device, respectively), which under ideal conditions will remove an equal amount of hydroxide or polyvalent cations from the sample. In practice, this maximum capacity may not be realized. The nature of the sample (concentration, ionic strength, solvent) and the rate at which the sample is loaded into the device will affect the capacity. Where possible, adjust sample size to use 50% or less of the device's absolute capacity. Larger volumes of hydroxide or polyvalent cations may be removed by using two or more Maxi-Clean™ cartridges in series. If the amount of hydroxide or polyvalent cations in the sample is unknown, a test extraction should be performed to determine the correct sample size.

For more information on the use of Maxi-Clean™ IC cartridges and the recovery obtained with each please see:

R. Saari-Nordhaus, J.M. Anderson, Jr. and I.K. Anderson, *Am. Lab.*, August (1990) 18.

I.K. Anderson, R. Saari-Nordhaus and J.M. Anderson, Jr., *J. of Chromatography*, 546 (1991) 61-71.

Sample Volume

The internal volumes of the different devices include the flow passages and interstitial packing volume. Although it is possible to recover all but 100µL (for the 0.5mL devices) or 150µL (for the 1.5mL devices) of sample with an air purge, best results are obtained when the sample volume greatly exceeds the internal volume of the cartridge.

Device	Bed Size	Internal Volume
® Maxi-Clean™ :	0.5mL	300µL
® Maxi-Clean™ Plus:	1.5mL	650µL
® Extract-Clean™ :	0.5mL	2.5mL
® Extract-Clean™ Plus:	1.5mL	1.5mL

Example

Trace Anions in Caustic

Procedure:

1. Apply 4mL of sample to a preconditioned Cartridge.
2. Discard the first 1mL of eluate.
3. Collect and analyze remaining eluate.

General Procedure



1. **Precondition the Device.** Pass 5 to 10mL of IC grade water through the device. This removes interstitial contaminants and wets the packed bed. For trace analysis work, repeat the rinsing procedure until the eluant is free from interferences.
2. **Load the Sample.** Load the entire sample at 1mL/minute or less. The total amount of hydroxide or polyvalent cations contained in the sample should not exceed the device capacity and preferably should be below 50% of total capacity. Discard the first 1mL of eluate. Collect the remaining eluate for analysis.

Other IC Devices

Device	Retains
IC-RP:	Hydrophobic Components
IC-OH:	Anions (pH increase)
IC-H:	Cations (pH reduction)
IC-Ag:	Chloride, Iodide, Bromide
IC-Ba:	Sulfate
IC-Na:	Cations (no pH change)
IC-Chelate:	Polyvalent Metal Ions
IC-Mixed Mode RP-OH:	Hydrophobic Components and Anions (pH increase)
IC-Mixed Mode RP-H:	Hydrophobic Components and Cations (pH reduction)

IC-H Devices

Description	Volume	Qty	Part No.
Maxi-Clean™ IC-H	0.5mL	50	5122575
Maxi-Clean™ IC-H Plus	1.5mL	25	5122568
Extract-Clean™ IC-H	0.5mL	50	5122910
Extract-Clean™ IC-H Plus	1.5mL	30	5122034



www.sstarpure.com

AUTHORIZED DISTRIBUTOR

MZ-Analysentechnik GmbH, Barcelona-Allee 17 • D-55129 Mainz

Tel +49 6131 880 96-0, Fax +49 6131 880 96-20

e-mail: info@mz-at.de, www.mz-at.de

S*Pure Pte Ltd, Singapore

Telephone: +65 656 339 01 Fax: +65 656 339 31 E-mail: corporate@sstarpure.com

EXTRACT-CLEAN™ and MAXI-CLEAN™ are trademarks owned by S*Pure Pte. Ltd.

The information presented herein is derived from our testing and experience. It is offered for your consideration and verification. Since operating conditions vary significantly, and are not under our control, we disclaim all warranties on the results that may be obtained from the use of our products. S*Pure Pte. Ltd. cannot be held responsible for any damage or injury occurring as a result of improper installation or use of its products. S*Pure reserves the right to change prices and/or specifications without prior notification.

S*PURE
www.sstarpure.com