DATA CODE: T00779

Application Title:
Separation of tryptic digest of beta-lactoglobulin

Compound:
Tryptic digest of beta-lactoglobulin



AUTHORIZED DISTRIBUTOR
MZ-Analysentechnik GmbH, Barcelona-Allee 17 • D-55129 Mainz
Tel +49 6131 880 96-0, Fax +49 6131 880 96-20
e-mail: info@mz-at.de, www.mz-at.de
MZ ANALYSENTECHNIK

Chromatographic conditions

Column:	TSKgel ODS-80TS , P/N 18150, 5 µm, 2.0 mm ID × 15 cm × 1
Mobile Phase:	A: 0.1% TFA B: 0.1% TFA in acetonitrile Linear gradient from A to 70% B for 30 min
Flow Rate:	0.2 mL/min
Detection:	UV/VIS @ 215 nm
Temperature:	25 °C
Injection Volume:	10 µL
Sample(s):	Tryptic digest of beta-lactoglobulin
Reference:	TOSOH Separation Report
Title:	Introduction of TSK-GEL semi-micro columns
Number:	97
Page start:	10
Page end:	10(2)
Comments:	UV (micro flow cell)