



Im Leuschnerpark 4,
64347 Griesheim (Darmstadt), Germany
Tel: +49 6155-70437-00
Fax: +49 6155-8357900
E-Mail: info.tbq@tosoh.com
Web: www.tosohbioscience.de

3604 Horizon Drive, Suite 100
King of Prussia, PA 19406, USA
Tel: +1 800-366-4875
Fax: +1 610-272-3028
E-Mail: info.tbl@tosoh.com
Web: www.tosohbioscience.com

OPERATING CONDITIONS and SPECIFICATIONS

TSKgel® Amide-80 Products

Part Numbers:**2 µm, UHPLC Compatible**

23454	2.0 mm ID x 5.0 cm L	2 µm	23460	Guard Column for 2 µm columns	
23455	2.0 mm ID x 10.0 cm L	2 µm			
23456	2.0 mm ID x 15.0 cm L	2 µm			
23457	3.0 mm ID x 5.0 cm L	2 µm			
23458	3.0 mm ID x 10.0 cm L	2 µm			
23459	3.0 mm ID x 15.0 cm L	2 µm			
20009	1.0 mm ID x 5.0 cm L	5 µm			
20010	1.0 mm ID x 10.0 cm L	5 µm			
21486	1.0 mm ID x 15.0 cm L	5 µm			
21487	1.0 mm ID x 25.0 cm L	5 µm			
21864	2.0 mm ID x 5.0 cm L	3 µm	21862	Guard Cartridge 2.0 mm ID x 1.0 cm L 3pcs	3 µm
21865	2.0 mm ID x 15.0 cm L	3 µm	19308	Guard Cartridge holder; for 2.0 mm ID columns	
19694	2.0 mm ID x 5.0 cm L	5 µm	21941	Guard column 2.0 mm ID x 1.0 cm L 3pcs	5 µm
19695	2.0 mm ID x 10.0 cm L	5 µm			
19696	2.0 mm ID x 15.0 cm L	5 µm			
19697	2.0 mm ID x 25.0 cm L	5 µm			
22850	3.0 mm ID x 5.0 cm L	3 µm			
22851	3.0 mm ID x 10.0 cm L	3 µm			
22852	3.0 mm ID x 15.0 cm L	3 µm			
21866	4.6 mm ID x 5.0 cm L	3 µm	21863	Guard Cartridge 3.2 mm ID x 1.5 cm L 3pcs	3 µm
22849	4.6 mm ID x 10.0 cm L	3 µm	19018	Guard Cartridge holder; for 3,2 mm ID columns	
21867	4.6 mm ID x 15.0 cm L	3 µm	19010	Guard column 3.2 mm ID x 1.0 cm L	5 µm
19532	4.6 mm ID x 5.0 cm L	5 µm	19021	Guard column 4.6 mm ID x 1.0 cm L	5 µm
19533	4.6 mm ID x 10.0 cm L	5 µm			
13071	4.6 mm ID x 25.0 cm L	5 µm	HR	High Resolution	
21982	4.6 mm ID x 25.0 cm L	5 µm			
14459	7.8 mm ID x 30.0 cm L	10 µm			
14460	21.5 mm ID x 30.0 cm L	10 µm	14461	Guard column 21.5 mm ID x 7.5 cm L	10 µm

This sheet contains the recommended operating conditions and the specifications for TSK-GEL Amide-80 columns. Installation instructions and column care information are described in a separate Instruction Manual.

A. OPERATING CONDITIONS

1. Shipping Solvent:

85 % Acetonitrile - 15 % Water	for 2µm, 3µm & HR columns	
75 % Acetonitrile - 25 % Water	for 5 µm & 10 µm columns, guardgel for 5 µm, 2 mm ID columns & guard column for 10 µm columns	
50% Acetonitrile/50% Water	for guardgel for 3 µm columns & guardgel for 5 µm, 4.6 mm ID column	
100% Acetonitrile	for guard column for 2 µm columns	

2. Max. Pressure:

40 kg / cm ²		2.0 mm ID x 5.0 cm L	2 µm
60 kg / cm ²		2.0 mm ID x 10.0 cm L	2 µm
80 kg / cm ²		2.0 mm ID x 15.0 cm L	2 µm
40 kg / cm ²		3.0 mm ID x 5.0 cm L	2 µm
60 kg / cm ²		3.0 mm ID x 10.0 cm L	2 µm
80 kg / cm ²		3.0 mm ID x 15.0 cm L	2 µm
30 kg / cm ²	= 427 psi	1.0 mm ID x 5.0 cm L	5 µm
60 kg / cm ²	= 853 psi	1.0 mm ID x 10.0 cm L	5 µm
120 kg / cm ²	= 1707 psi	1.0 mm ID x 25.0 cm L	5 µm
90 kg / cm ²	= 1280 psi	10.0 mm ID x 5.0 cm L	5 µm
40 kg / cm ²	= 569 psi	2.0 mm ID x 5.0 cm L	5 µm
80 kg / cm ²	= 1138 psi	2.0 mm ID x 10.0 cm L	5 µm
100 kg / cm ²	= 1422 psi	2.0 mm ID x 15.0 cm L	5 µm
150 kg / cm ²	= 2134 psi	2.0 mm ID x 25.0 cm L	5 µm
150 kg / cm ²	= 2134 psi	4.6 mm ID all columns	5 µm
70 kg / cm ²	= 996 psi	7.8 mm ID x 30.0 cm L	10 µm
30 kg / cm ²	= 427 psi	21.5 mm ID x 30.0 cm L	10 µm
200 kg / cm ²	= 2845 psi	For all 3 µm columns	

3. pH Range: 2.0 - 7.5

4. Organic Conc.:

0 - 100%	polar and nonpolar	5µm, including HR & 10µm switch to 100% acetonitrile before switching to 100% nonpolar solvents.
0 - 100%	polar and nonpolar	2µm & 3 µm: switch to 95% acetonitrile before switching to 100% nonpolar solvents.

5. Temperature:

10°C - 80°C	5 µm and 10 µm columns; including HR
10°C - 50°C	2 µm and 3 µm columns;

6. Cleaning Solvents:
 - (1) Water
 - (2) 45% Acetonitrile or acetone
 - (3) 0.1% triethylamine in at least 75% acetonitrile
 - (4) 50 mM phosphate buffer pH 6.0 / Acetonitrile 50/50

NOTE: Choose cleaning solvent (1) to remove very polar solutes from the bonded phase, (2) to desorb hydrophobic compounds, (3) to get rid of adsorbed basic compounds, and (4) for ionic compounds.

7. Storage: Store the column in the shipping solvent when it will not be used the next day. For overnight storage flush the column with the mobile phase at 50% of the recommended flow rate for the column.

8. Column Protection: The use of guard columns is recommended to prolong the life of the analytical column. Guard column life depends greatly on sample cleanliness. As a general rule, guard columns should be replaced after every 30-40 sample injections, when the peaks become excessively wide, or when the peaks show splitting.

B. SPECIFICATIONS


The performance of TSK-GEL Amide-80 columns is tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications:

No. of Theoretical Plates (N)	5,800	2.0 mm ID x 5.0 cm L	2µm	8,300	3.0 mm ID x 5.0 cm L	2µm
	14,000	2.0 mm ID x 10.0 cm L	2µm	16,500	3.0 mm ID x 10.0 cm L	2µm
	21,500	2.0 mm ID x 15.0 cm L	2µm	24,000	3.0 mm ID x 15.0 cm L	2µm
	≥ 300	1.0 mm ID x 5.0 cm L	5 µm	≥ 6,000	4.6 mm ID x 5.0 cm L	3 µm
	≥ 600	1.0 mm ID x 10.0 cm L	5 µm	≥ 12,000	4.6 mm ID x 10.0 cm L	3 µm
	≥ 4,000	1.0 mm ID x 15.0 cm L	5 µm	≥ 18,500	4.6 mm ID x 15.0 cm L	3 µm
	≥ 6,000	1.0 mm ID x 25.0 cm L	5 µm	≥ 13,000	4.6 mm ID x 25.0 cm L	3 µm
	≥ 1,000	2.0 mm ID x 5.0 cm L	5 µm	≥ 1,500	4.6 mm ID x 5.0 cm L	5 µm
	≥ 2,000	2.0 mm ID x 10.0 cm L	5 µm	≥ 3,000	4.6 mm ID x 10.0 cm L	5 µm
	≥ 4,000	2.0 mm ID x 15.0 cm L	5 µm	≥ 8,000	4.6 mm ID x 25.0 cm L	5 µm
	≥ 6,000	2.0 mm ID x 25.0 cm L	5 µm	≥ 18,000	4.6 mm ID x 25.0 cm L	5 µm HR
	≥ 3,500	2.0 mm ID x 5.0 cm L	3 µm			
	≥ 13,000	2.0 mm ID x 15.0 cm L	3 µm	≥ 5,000	7.8 mm ID x 30.0 cm L	10 µm
	≥ 5,500	3.0 mm ID x 5.0 cm L	3 µm	≥ 8,000	21.5 mm ID x 30.0 cm L	10 µm
	≥ 11,000	3.0 mm ID x 10.0 cm L	3 µm			
	≥ 16,000	3.0 mm ID x 15.0 cm L	3 µm			

Please note, that there are not all theoretical plate specifications for 2 mm ID columns available.

Asymmetry Factor (AF)	All 2µm columns					
	0.90 - 1.60	1.0 mm ID x 5.0 cm L	5 µm	0.75 - 1.40	4.6 mm ID x 5.0 cm L	5 µm
1.00 - 2.00	1.0 mm ID x 10.0 cm L	5 µm	0.70 - 1.50	4.6 mm ID x 10.0 cm L	5 µm	
0.70 - 1.80	1.0 mm ID x 15.0 cm L	5 µm	0.70 - 1.60	4.6 mm ID x 25.0 cm L	5 µm	
0.70 - 1.80	1.0 mm ID x 25.0 cm L	5 µm	0.90 - 1.30	4.6 mm ID x 25.0 cm L	5 µm HR	
0.90 - 1.25	2.0 mm ID x 5.0 cm L	3 µm	0.90 - 1.20	4.6 mm ID x 5.0 cm L	3 µm	
0.90 - 1.25	2.0 mm ID x 15.0 cm L	3 µm	0.90 - 1.20	4.6 mm ID x 10.0 cm L	3 µm	
0.70 - 1.80	2.0 mm ID x 15.0 cm L	5 µm	0.90 - 1.20	4.6 mm ID x 15.0 cm L	3 µm	
0.70 - 1.80	2.0 mm ID x 25.0 cm L	5 µm				
0.90 - 1.25	3.0 mm ID x 5.0 cm L	3 µm				
0.90 - 1.25	3.0 mm ID x 10.0 cm L	3 µm				
0.90 - 1.25	3.0 mm ID x 15.0 cm L	3 µm				

Please note, that there are not all asymmetry specifications for 2 mm ID (5 µm), 7.8 mm ID (10 µm), and 21.5 mm ID columns available.



AUTHORIZED DISTRIBUTOR

MZ-Analysentechnik GmbH, Barcelona-Allee 17• D-55129 Mainz
 Tel +49 6131 880 96-0, Fax +49 6131 880 96-20
 e-mail: info@mz-at.de, www.mz-at.de